

**OKLAHOMA STATE BUREAU OF INVESTIGATION
CRIMINALISTIC SERVICES DIVISION**

Controlled Substances Quality Assurance Manual
Effective 07-20-18

ATTENTION: If any portion(s) of the Controlled Substances Quality Assurance Manual or any of the Controlled Substances Policies is/are unclear to any analyst or if a circumstance arises outside the scope of these documents, it is the responsibility of each analyst to immediately notify the Technical Manager and the respective Supervisor to seek clarification/approval or obtain guidance on the issue BEFORE proceeding.

This manual is set out in several sections that deal with different areas relating to the quality of the work performed by the Criminalists in the controlled substance units.

Table of Contents

Objective.....	1
1. Equipment Maintenance/Calibration.....	2
2. Reagents.....	4
3. Solvents.....	6
4. Drug Reference Standards.....	8
5. Quick Reference Guide for Labeling Requirements in the Controlled Substances Laboratory.....	11
6. Weight Reference Standards.....	12
7. Case Management and Analysis.....	13
8. Abbreviation List.....	18
9. Standard Forms for Documentation.....	19
10. Documentation for the Training Manual.....	19
11. Examination Documentation.....	20
12. Case Reviews.....	21
Approval.....	23
History.....	24

Objective

Throughout the Criminalistics Services Division of the OSBI, quality in all aspects of the work performed is sought. It is the goal of all Criminalists involved in the identification of controlled substances to provide the most accurate analyses possible. In order to ensure that this goal is being attained throughout the system of controlled substance identification units, this quality manual has been established. Some issues that concern specifically controlled substance analysis not covered in division policy will be addressed in this manual. The purpose of the Controlled Substances Quality Assurance Manual is to ensure consistency of practices among all of the controlled substance laboratories within the OSBI system and to ensure that the quality of casework is maintained.

Changes to the Drug Identification Quality Manual:

Corrections, additions, or deletions to the Controlled Substances Quality Assurance Manual should be put forth to the Technical Manager in writing. This can be done by any analyst within any of the controlled substance identification units of the OSBI. If it is determined that a change is necessary then the current procedure QP3, Deviations, in the laboratory quality manual will be followed.

1. Equipment Maintenance/Calibration

1.1 Purpose:

Equipment maintenance and maintenance documentation is essential for the smooth operation of a laboratory. By accurately documenting maintenance performed on laboratory equipment, a record of quality assurance is provided that can be referenced, if necessary, to troubleshoot future problems or to show that the highest level of integrity possible in casework results is being produced. It is the purpose of this protocol to establish a set of guidelines that are specific for the maintenance of equipment that will be utilized in the analysis of controlled substances.

1.2 Policy:

- A. Supervisors and analysts will be responsible for assuring that all maintenance on equipment involved in the identification of controlled substances is being performed and properly documented.
- B. All criminalists within the controlled substance discipline will be able to recognize when equipment maintenance is necessary and are charged with maintaining laboratory equipment and documenting all maintenance performed.

1.3 Definitions:

Maintenance Log: Documentation that is kept detailing maintenance that is performed on laboratory equipment.

In Routine Use: Equipment that is used and maintained on a regular basis.

Not in Routine Use: Refers to equipment that is in working order but is not used routinely enough to have regular entries in its maintenance log. Some instruments not in routine use may require some maintenance performed prior to use.

Out of Service: Equipment that cannot be used for casework. These pieces of equipment may not be operational and may require servicing and/or calibration prior to being placed back into service.

1.4 Procedure:

- A. Maintenance logs will be kept on equipment that is used in the analysis of controlled substances. All maintenance that is performed will be documented in the maintenance log sheet for each piece of equipment. The maintenance log will contain the identifiers for the instrument including any name given to the instrument and its OSBI asset number. The log will include an explanation of why maintenance was performed, what maintenance was performed, the date maintenance was performed, and the initials of the analyst performing the maintenance. Maintenance performed by someone outside of the unit (i.e. ASAP Analytical, Agilent Technologies, etc.) will also need to be documented on the maintenance log. If parts of, or the entire instrument need to be sent away for repairs this will be noted in the maintenance log as well. Replacement or repaired equipment will be function tested by running the default sequence consisting of a positive and negative control and documented in the maintenance log prior to being used for casework. DRQM-1 is a General Maintenance Log; instrument maintenance logs are incorporated into the Controlled Substances

Protocol Manual.

- B. Maintenance that is performed on a regular basis such as monthly or when count limits are met (i.e. the changing of liners for the GC and GC/MS or hydrocarbon ladder on the GC/MS) needs to be noted in the maintenance log. No explanation for the action is necessary unless the maintenance is being performed for a reason other than routine maintenance (i.e. liner changed because of carryover). Counts for liner, septa, and gold seal changes should be documented in the maintenance log when changed. Liner/Septa changes are considered routine at a count of 225 or higher. *Changing only the glass wool and not changing the liner/septa is not permitted.* Gold seal changes are considered routine at 1200 for the GC and 1450 for the GCMS. Split vent line changes are considered routine at 4800 or higher. Split vent filter changes are considered routine at 2400 or higher for the GC/MS and 4800 or higher for the Agilent 7890 GC.
- C. It is not necessary to keep the maintenance log for each piece of equipment needing to be maintained in a separate book. Instrument sets (GC and GCMS connected to the same data station) and miscellaneous pieces of equipment, such as hydrogen generators, that have maintenance logs can be kept in one binder but in separate sections. Individual instruments such as a GC, GC/MS, FTIR, or GC/IRD should each have their own maintenance log. Instrument/equipment maintenance logs can be kept in electronic form; if kept electronically, each instrument will have its own folder.
- D. If a piece of equipment requires maintenance that cannot be performed right away, then it should be labeled “Out of Service” and should not be used. This should also be documented in the maintenance log with an explanation of why the instrument was removed from service along with the date and initials of the analyst removing it from service. Once the proper maintenance has been performed and the instrument performs as expected, it can be returned to service and the maintenance documented.
- E. If a piece of equipment is not in routine service, it should be labeled as “Instrument not in Routine Use, Performance Checks Must Be Performed Prior to Laboratory Analysis.” This should also be reflected in the maintenance log.
- F. Items that should have maintenance logs include:
 - Gas Chromatographs (DR30-1)
 - GC/IRD (DR75-1)
 - FTIR (DR60-1)
 - Gas Chromatograph/Mass Spectrometers (DR70-1)
 - Hydrogen Generators (DRQM-1)
 - Air Compressors (if applicable) (DRQM-1)

Individual balances/scales will have forms (DR4-1, DR4-2, & DR4-3) for calibration checks and documentation of calibration checks and maintenance. Balances/scales are the only equipment that requires a calibration, see protocols for calibration requirements.

G. A schedule of routine maintenance to be performed in the laboratory is as follows:

Maintenance:	When performed:	Protocol Reference:
Glass liners/septa changed	At or before <u>250</u> injections	DR-30, DR-70 and DR-75
Gold seal changed	At or before <u>1250</u> injections for GC; at or before <u>1500</u> injections for GC/MS and GC/IRD	DR-30, DR-70 and DR-75
Clean FID jets for each GC	When the gold seal is changed	Quality Manual 1.4-G
Standard ladder run	First five working days of the month, prior to use of a GC after a new column is installed, or every time the column is cut.	DR-30
Hydrocarbon Ladder	First five working days of the month, prior to use of a new column for the GC/MS, or every time the column is cut.	DR-70
Cocaine standard	Prior to the analysis of casework each day on both the GC and GC/MS. See DR-70 for <i>continuous</i> sequence exception.	DR-30, DR-70 and DR-75
Blanks	On the GC a solvent blank needs to be run every fifth sample (four samples then a blank). On the GC/MS and GC/IRD a blank made up of the solvent or extraction that is used for the samples in that case needs to be run during the samples of each new case and at least every sixth injection (five samples and a blank) within a case.	DR-30, DR-70 and DR-75
Balance calibration check	Daily (500g) and Monthly	DR-4
Deionizer bag replacement on hydrogen generators	When change water light comes on or every 6 months, whichever comes first.	Quality Manual 1.4-G
Drain water for air compressors	Unless equipped with an auto drain, at least monthly	Quality Manual 1.4-G
Desiccant change for air compressors	When visual inspection of desiccant indicates a change is necessary	Quality Manual 1.4-G
Desiccant change for hydrogen generators	When visual inspection of desiccant indicates a change is necessary	Quality Manual 1.4-G
Split vent line for GC and GCMS changed	At or before <u>5000</u> injections	Quality Manual 1.4-G
Split vent filter for GCMS changed	At or before <u>2500</u> injections	Quality Manual 1.4-G
Split vent filter for Agilent 7890 GC changed	At or before <u>5000</u> injections	Quality Manual 1.4-G

H. Refrigerators should have a Temperature Monitoring Form that is maintained. Refer to OSBI CSD QP 6.4 for form and a monitoring schedule.

2. Reagents

2.1 Purpose:

The use of reagents in casework is a common practice throughout the controlled substance laboratory system of the OSBI. Most of these reagents are made by the Criminalists working in the controlled substance laboratories through the combination of the proper chemical components. The reagents used must be checked to ensure that they function as expected and documented to ensure quality and accountability.

2.2 Policy:

A. All Criminalists in the controlled substance identification units will know how to properly prepare and QA/QC all reagents that are used in the identification of controlled substances.

- B. Supervisors and analysts will be responsible for verifying that historical reagent formulation sheets are properly archived and reagents are properly logged into ChemInventory, and that function verifications are being performed and documented.

2.3 Definitions:

Reagent: A substance used in a chemical reaction to detect, measure, examine, or produce other substances.

Function Verification: A test that is performed on a reagent to verify that it will function as expected.

2.4 Procedure:

- A. When reagents are made, the following will be documented in ChemInventory:
1. Reagent Name
 2. Lot Number
 3. Expiration (if applicable)
 4. Components, including:
 - i. Quantity of each component
 - ii. Supplier of each component
 - iii. Lot Number of each component
 - iv. Expiration Date of each component (if applicable)
 5. Steps for preparation, including:
 - i. Analyst name preparing the reagent
 - ii. Date prepared
 6. Reagent function verification procedure, including
 - i. If the Lot **DOES** function as expected, the analyst will log their name and date in the “QC by” and “QC date” fields in Chemical Inventory
 - ii. If the Lot **DOES NOT** function as expected, the analyst will document this in the “Notes” for the reagent in Chemical Inventory along with:
 1. Analyst name performing the verification procedure
 2. Date verification procedure performed

After a reagent lot has been entered into ChemInventory, if the lot is moved to a regional lab it does not need to be re-entered into ChemInventory.

- B. Prepared reagents will be labeled with:
1. the identity of the reagent
 2. the lot number. The lot number assigned will be the date the reagent was made; if more than one lot is made on the same day the additional lots will be designated by a dash and a sequential number (i.e. Lot# 103116-2)
 3. the initials of the analyst that prepared the reagent
 4. storage requirements (as applicable)
 5. If the reagent is transferred into another container, the container will be labeled with the

identity of the reagent, the lot number, the initials of the analyst transferring the reagent, and storage requirements (as applicable).

- C. Reagents such as the standard ladder and hydrocarbon ladder will be made in the FSC Laboratory and distributed to the regional laboratories. If used, copies of the formulation sheets will be uploaded into ChemInventory, but do not need to be distributed with the reagent.
- D. All chemicals used to make reagents will have the name of the chemical, the manufacturer's name, lot number, date received, initials of the analyst that received the chemical, date opened, and initials of the analyst who opened the container on the chemical container.
- E. At a minimum, the reagent will have a function verification before being used for casework. Reliability/verification may also occur concurrent with casework, as a part of the quality control system. The individual components of a reagent do not have to be verified with the exception of the standard ladder.
- F. All reagents routinely used in the course of casework will have the lot number recorded in the Reagent Information panel for each case within the BEAST. Reagents occasionally used in casework that are not currently listed in the Reagent Information panel shall have the lot numbers recorded in the analyst notes.

3. Solvents

3.1 Purpose:

Solvents play an essential part in the analysis of controlled substances throughout the system of controlled substance laboratories. They are used in a variety of ways that include dissolving casework samples and making reagents. The handling of solvents must ensure that they are kept free from contaminants to maintain quality in laboratory practices.

3.2 Policy:

- A. Analysts will be responsible for verifying that solvents are free from contamination and will be able to recognize when possible contamination of a solvent exists. Analysts will also know what actions to take in the event of possible solvent contamination.
- B. Supervisors and analysts will be responsible for verifying historical solvent logbooks are properly archived and solvents are properly logged into ChemInventory, and that quality verifications are being performed and proper documentation stored in ChemInventory.

3.3 Definitions:

Solvent: A substance that is intended to provide a solution.

3.4 Procedure:

- A. When a new solvent bottle is opened, the analyst opening the bottle will write his or her initials on the bottle and the date the bottle was opened.

- B. In most cases solvents should be ordered and shipped to the laboratory where they will be used. In some instances, it may be necessary to transport solvents between laboratories in the OSBI system. This can be done by either transporting the solvent in its original shipping container or by placing a solvent bottle in a bucket with a lid and filled with vermiculite. Great care should always be taken when transporting solvents.
- C. All solvents in their original containers will bear:
1. Solvent name
 2. Lot number
 3. Manufacturer
 4. Date and initials of the analyst receiving the container
 5. Date and initials of the analyst opening the container.
 6. If the solvent is transferred into another container, the new container will be labeled with the name of the solvent, the lot number, and the initials of the person transferring the solvent.
- D. When solvents are received, the following will be documented in ChemInventory:
1. Solvent name
 2. Lot number
 3. Manufacturer
 4. Date and initials of the analyst receiving the container
 5. Date and initials of the analyst opening the container.
 6. Solvent function verification procedure, including
 - i. If the Lot **DOES** function as expected, the analyst will log their name and date in the “QC by” and “QC date” fields in Chemical Inventory
 - ii. If the Lot **DOES NOT** function as expected, the analyst will document this in the “Notes” for the reagent in Chemical Inventory along with:
 1. Analyst name performing the verification procedure
 2. Date verification procedure performed
- E. Before a new lot of a solvent can be used, it needs to be shown that it is free of contaminants. This can be done by taking a sample of the solvent and analyzing it on the GCMS. The GCMS scan will be archived in ChemInventory. If the GCMS analysis shows no sign of contamination, the solvent lot can be used. If the solvent appears to be contaminated, then the supervisor and Technical Manager should be notified, and the solvent lot will not be used. If used, copies of the solvent log will be uploaded into ChemInventory, but do not need to be distributed with the solvent.
- F. All solvents routinely used in the course of casework will have the lot number recorded in the Reagent Information panel for each case within the BEAST. Solvents occasionally used in casework that are not currently listed in the Reagent Information panel shall have the lot numbers recorded in the analyst notes.

4. Drug Reference Standards

- A. Drug reference standards are used throughout the system of controlled substance laboratories within the OSBI for a variety of purposes. They are used for comparison in casework, verification of proper instrument operation, method validation, and research. The handling of standards must ensure that they remain free from any contamination, and each standard must be verified to ensure its identity. The receipt and usage of all drug reference standards (controlled and non-controlled) will be documented using this procedure.
- B. Drug standards are considered reference materials.
- C. Each laboratory site of the OSBI will be properly permitted by the Oklahoma State Bureau of Narcotics and Dangerous Drugs Control (OBNDDC), and by the Federal Drug Enforcement Administration (DEA) when controlled substances are stored at the site. It will be the responsibility of each regional Criminalist Supervisor and the Criminalistics Division Director in the Forensic Science Center, FSC, to ensure their permits remain current.
- D. All standards will be ordered, shipped, and logged into the FSC located in Edmond.
- E. All drug reference standards that are controlled under Title 63, O.S. Sections 2-204, 2-206, 2-208, 2-210, and 2-212 should be maintained in alphabetical order in an appropriated secure location. This location can include a safe, a locked drawer, a refrigerator that is maintained in a secure area, or any other area where access is limited to appropriate personnel. The Criminalist Supervisor assigned to that particular laboratory or section will have the discretion to designate the Criminalist(s) that will have access to the drug standards.
- F. The receipt and usage of all drug standards will be accounted for on a Drug Reference Standard Receipt and Usage Log (DRQM-4). The log form will also be filled out completely unless a specific notation is made as to why information is being left out. Verification information may be left blank until the verification is performed. A separate reference standard log form will be used for each container except for DEA exempted analytical reference standards where one form can be used for multiple containers of the same lot number. If there is more than one container with the same lot number, each container will be numbered (1, 2, 3, etc.) and a separate log sheet made for each with the container number listed on the log form. If more than one log form is required to document the usage of a standard before it is expended, additional forms will be created as needed and the forms numbered. If a mistake is made on a log form, it is to be corrected by a single line strike out with initials and the correct information then written. Do not make obliterations and do not use whiteout on forms.
- G. When the drug reference standard is opened, the initials of the analyst opening the container and the date opened will be documented on the container.
- H. Before a drug reference standard from a new lot can be used, its identity must be verified by a GC/MS, GC/IRD, or FTIR. This is done by simply analyzing the standard by GC/MS, GC/IRD, or FTIR and comparing the scan to a literary reference. The source of the literary reference can be a

book, journal, previously established library entry such as GBI or NBS, or an independent analysis.

1. If a drug reference standard cannot be verified by instrumental analysis, due to reasons other than a lack of reference data, then the Technical Manager must be notified. The supplier may be removed from the list of approved drug reference suppliers and the drug reference standard will not be used in casework.
 2. All scans will be maintained in the standard logbooks, or in the ContSub\Standard_Verifications folder on the OSBI network, with the original log forms. The method of verification, source of the literary comparison, and the name and date of the analyst performing the verification will also be documented in the standard logbook or in the ContSub\Standard_Verficiation folder.
 3. In some instances, standards will be labeled as the dextro, levo, or dextro/levo rotary isomer. If the compound is controlled under Title 63 of the Oklahoma State statutes with wording that would include both isomers to be controlled as the same substance (example: Methamphetamine), then it is not necessary verify the substance as dextro, levo, or dextro/levo rotary. If the substance is controlled as the dextro, levo, or dextro/levo rotary isomer (example: Levomethorphan), then it is necessary to verify the dextro, levo, or dextro/levo rotary isomer.
- I. A secondary standard is a solution with a known concentration prepared from a reference standard. When a secondary standard is made (i.e. 30 mg of methamphetamine diluted in methanol), a chemical formulation sheet is not required. The lot number reflected on the container will be the lot number of the standard. Along with the lot number, the container will be labeled with the standard name, the solvent system used, the lot number of the solvent, the initials of the analyst making the secondary standard, and the date made. If autosampler vials containing secondary standards are kept/stored, the vial must be labeled with ALL the information required on secondary standards.
- J. If a drug standard is to be taken to a regional laboratory, it will be logged out under the regional laboratory's DEA number and a photocopy made of the original form and all verification data for that standard. A note will be made on the original form documenting to which regional laboratory the standard has been transferred and the person transferring it. The photocopied form and verification data will then be taken to the regional laboratory and placed into a log book of current drug reference standards. The originals will always remain at the FSC. The amount taken will be entered as the total amount remaining on the copy taken to the regional laboratory and the usage documented from that point. In these instances, all identifiers (drug name, lot number, manufacturer, etc.) are to be placed on the container utilized to transfer the drug to the region unless the entire original container is taken. If the entire container is to be transferred, a note will be made on the original log and placed with its forms of verification in the expended drug standards log book or in the ContSub\Expended_Standards folder on the OSBI network. If a standard is to be verified in a regional laboratory after transfer from the FSC, then a notation will be made on the original that the verification will be done in the regional laboratory.
- K. Drug standards that need to be destroyed are to be zeroed out and a note will be made indicating that the drug has been destroyed on the reference standard log form. The log form will then be placed in

- the expended drug standards notebook or scanned and stored in the ContSub\Expended_Standards folder on the OSBI network. A drug destruction inventory form will then be filled out and submitted to the Admin. Program Officer for Physical Evidence with the drug for destruction as required under Title 63, O.S. Section 2-315. A copy of the destruction form is to be maintained in a permanent file maintained by the Administrative Program Officer for Physical Evidence.
- L. Standard verification data and the initial drug log sheet will be kept in alphabetical order in a log book of current drug reference standards until expended, or will be stored in the ContSub\Standard_Verifications folder on the OSBI network.
 - M. Drug log sheets with inventories pending will be kept in alphabetical order in a log book of current drug reference standards. This log book will be kept in the unit responsible for maintaining the standards.
 - N. When a drug inventory is depleted, the log sheet will be zeroed out in the amount remaining column and the form maintained in alphabetical order in a log book of expended drug standards or scanned and stored in the ContSub\Expended_Standards folder on the OSBI network. This log book will be kept in the unit responsible for maintaining the standards.
 - O. Under no circumstances will standards be transferred out of the OSBI Criminalistics Laboratory System without the authorization of the designated Criminalistics Administrator and the DEA forms being completed.
 - P. It will be the responsibility of each authorized individual handling drug standards as outlined in #5 of this section to ensure compliance with this policy. Violations of this policy may result in disciplinary action and possible criminal charges under the Uniform Controlled Dangerous Substances Act.
 - Q. It is understood that the OSBI keeps a daily “working” inventory of drug standards. When secondary standards are created from primary standards, the secondary standard will be analyzed by the GC/MS, GC/IRD or FTIR to ensure it has not been altered. If additional aliquots of secondary standard are created, the instrumental data will be kept with the log sheet. Once the primary standard has been expended, all additional instrumental data will be stored with the log sheet in the log book for expended standards or scanned and stored in the ContSub\Expended_Standards folder on the OSBI network.

5. Quick Reference Guide for Labeling Requirements in the Controlled Substances Laboratory

Below are the items that are to be labeled and the labeling that needs to appear on each item in the laboratory. Also included is the reference that directs such action.

Standards	<p>Reference Standard containers will be labeled with:</p> <ul style="list-style-type: none"> The initials of the analyst who received the standard and date received. The initials of the analyst who opened the container and date opened, manufacturer, name of standard, and lot number. <p>All secondary standards will have:</p> <ul style="list-style-type: none"> The name of the standard, lot number of the standard, initials of the analyst who made the secondary standard, date secondary standard was made, solvent, and solvent lot number. 	CSD QP26; Quality Manual 4.7 & 4.9
Reagents	<p>Upon receipt of all reagents, the initials of the analyst who received and the date received will be marked on the container. On commercially prepared reagents (Hydrogen Peroxide), the manufacturer's name, the initials of the analyst, and date opened will be marked on the container.</p> <p>All reagents will have the name of the reagent, lot number, identity of analyst preparing it, expiration date (if applicable) and storage requirements (as applicable)</p>	CSD QP-08; Quality Manual 2.4-D
Solvents in original manufacturer's container	These solvent containers will bear the name of the solvent, lot number, manufacturer's name, date and initials of the analyst receiving the container, and the date and initials of the analyst opening the container.	Quality Manual 3.4-C
Solvents in other containers	These solvent containers will bear the name of the solvent, lot number, and the initials of analyst who transferred the solvent.	Quality Manual 3.4-C
All other chemicals without log sheets	The container should bear the name of the chemical, the manufacturer, lot number, date the container was received, initials of analyst opening the container, and date the container was opened.	Quality Manual 2.4-D

6. Weight Reference Standards

6.1 Purpose:

Weights are used monthly and daily to demonstrate that the scales and balances used to weigh case samples are working correctly. The proper storage, handling and transportation of weights ensure that they are kept free from contamination or deterioration and protect their integrity.

6.2 Policy:

- A. The certified reference standard weights used for balance/scale verification will be NIST Class F, ANSI/ASTM Class 6 or better, and will be calibrated and certified annually by an outside vendor that is accredited to ISO/IEC 17025:2005 by an accrediting body that is a signatory to the ILAC Mutual Recognition Arrangement, with the calibration to be performed listed in the scope of accreditation. These weights are viewed as critical and the vendor weights are purchased from shall be evaluated.
- B. The annual Certificate of Measurement Traceability, Certificate of Accreditation to ISO/IEC 17025, and Scope of Accreditation to ISO/IEC 17025 for the outside vendor will be maintained by the Forensic Chemistry Technical Manager.
- C. The certified reference standard weights will be kept in storage containers that will protect the weights from contamination and damage. The weights will be transported in the respective storage containers and not in buckets and/or boxes.
- D. The surrogate reference weights for the Quality Assurance Studies will be issued from FSC Forensic Chemistry Lab. The surrogate reference weights must be stored in a manner that ensures that they are kept free from contamination and damage. Any problems/issues with the surrogate weights will be reported to the technical manager and the weights shall not be modified or changed, without authorization from the technical manager.
- E. The reference standard weights will be used monthly to verify a scale is working correctly. If, at any time, a weight is outside of the acceptable 1% range on a scale, the weight will be weighed on a second scale to resolve that the issue is with the scale and not the weight. If it is determined the issue is with the weight, the weight will be sent to an outside vendor for re-calibration and re-certification before it can be used again. The outside vendor must be accredited to ISO/IEC 17025:2005 by an accrediting body that is a signatory to the ILAC Mutual Recognition Arrangement, with the calibration to be performed listed in the scope of accreditation.
- F. Weight is a reported test result and requires traceability and uncertainty. Uncertainty is only reported for cases involving trafficking weights. Approximate volume, is a description of evidence and does not require traceability or uncertainty.

7. Case Management and Analysis

7.1 Purpose:

It is necessary that casework and case files be handled similarly among all analysts. To ensure consistency throughout the system of controlled substance laboratories within the OSBI, this section will address issues regarding the analysis of casework samples and case reporting.

7.2 Policy:

- A. It is the responsibility of all analysts in the OSBI Controlled Substances Identification Units to ensure that accuracy and quality are maintained and that the necessary work required to maintain the integrity of results is put forth in all cases.
- B. The technical manager, supervisors, or their designees are charged with reviewing the performance of analysts through case review, proficiency testing, or whatever measures are deemed necessary to ensure that consistency and quality throughout the laboratory system is maintained.

7.3 Procedure:

- A. The following is a list of examinations available in the drug identification units:

<u>Confirmatory Tests (A)</u>	<u>Non-Confirmatory Tests (B)</u>	<u>Presumptive Tests (C)</u>
GC/MS	TLC	Marquis
FTIR	GC	Cobalt Thiocyanate
GC/IRD	Literary References	Bates
	Visual/Microscopic examination	Duquenois/Levine
		Ehrlich's (LSD)

For a conclusive analysis of a controlled dangerous substance (with the exception of marijuana), an analyst needs a positive presumptive or non-confirmatory test that agrees with a positive confirmatory test, or two confirmatory tests that agree. Before using the GC/MS or GC/IRD, the sample must first be analyzed on the GC. Marijuana requires a positive visual examination, TLC examination, and Duquenois/Levine examination, or a positive visual examination with a GC and GC/MS or GC/IRD that is positive for Tetrahydrocannabinol.

- B. In the analysis of pharmaceuticals, a literary reference is sometimes sufficient to presumptively report results as outlined in DR-5 of the OSBI Controlled Substances Protocol Manual. If a literary reference is found that matches the logo on the pharmaceutical and the analyst proceeds to analyze the substance to confirm its identity, then the analyst should conclusively identify the substance. If a substance cannot be confirmed, it is not acceptable to simply report "No controlled substances identified." If it is not possible to confirm the substance due to a situation like a weak GC/MS or poor FTIR, then the report wording should indicate that instrumental analysis and literary reference indicate that a controlled substance is present but was unable to be confirmed. Notes should also be made in the case file as to why the substance was not confirmed, so that when reviewed at a later date, the reviewer will understand why the substance was not confirmed.

- C. Reports relating to the analysis of tablets and capsules will reflect the number of tablets, tablet fragments or capsules analyzed and confirmed.
1. If a case contains multiple items of tablets, tablet fragments or capsules and one tablet, tablet fragments or capsule was analyzed for each item, one statement at the bottom of the result section will be acceptable for this requirement.
 2. If a case contains multiple tablets of one tablet/tablet fragments/capsules type (color, shape, logo, etc.), one statement at the bottom of the result section will be acceptable for this requirement.
 3. If an item(s) contains tablets, tablet fragments or capsules and more than one is analyzed, results will be reported as follows:
Item #: Controlled substance, Schedule XX (five tablets analyzed and confirmed)
or
Item #: Controlled substance, Schedule XX (five tablets analyzed, one tablet confirmed)
 4. If the case has single tablet item(s) and tablet(s) is/are tested, then it is up to the analyst to decide if they want to remove the statement from the report.

All tablet/capsule samples should consist of a maximum of ½ of the tablet, tablet fragment or capsule being sampled. If multiple tablets, tablet fragments or capsules are tested then only ½ of each tablet/capsule should be sampled and analyzed individually, the halves/fragments cannot be combined. The remaining tablet/tablet fragment/capsule halves should be marked and returned to evidence.

- D. For each item containing tablets/tablet fragments/capsules, it is necessary to label each item and what was analyzed, by either wrapping the item in tape or placing the item in an individual plastic bag and labeling. When ½ of a tablet is used for sampling, the other half should be labeled. It is important that all items are labeled clearly. For example, if one container contains four types of tablets, one of each type of tablet should be labeled with the item number, lab number, date, and initials of the analyst. If a container only contains one type of tablet, then it is acceptable to label the container with the item number, lab number, date, and initials of the analyst and wrap the tested tablet with tape and initial the tape.
- E. When analyzing blotter squares/gelatin squares/candy/individual or multiple dosage units, it is necessary to clearly indicate what was analyzed. This is accomplished with item labeling of the evidence, documentation in case file notes and notations on the report. The following uses “sheets” as the descriptor, but it may apply to evidence in forms other than sheets, i.e. pieces of candy.

Labeling:

If you have multiple similar sheets of evidence with dosage units, you may inventory them by the following methods:

1. Label all sheets of evidence with one item number and take a sample from ONE sheet and analyze it, and apply results to all (since they are considered one item).
2. Separate one sheet as one item and analyze it and report results. All other sheets are a different item with no analysis.
3. Label the sheets with different item numbers and take samples from different sheets, analyze them separately and report the individual results.

You may not label all sheets as one item, take multiple samples from different sheets and combine them into one sample for analysis.

If you have multiple single dosage units, you may inventory them by the following methods:

1. Make one dosage unit its own item and analyze $\frac{1}{2}$ of it. All remaining dosage units are a different item with no analysis.
2. Make multiple dosage units into one item, take $\frac{1}{2}$ of one square, analyze it, and apply results to all dosage units (since they are considered one item).

Sampling of dosage units:

If you have a piece of evidence with individual dosage units, i.e. may be separated with perforations, you may:

1. Take $\frac{1}{2}$ of one dosage unit for analysis
2. Take $\frac{1}{2}$ of many connected dosage units for analysis (2 or more)

Whichever labeling and sampling methods are chosen, it must be documented in the item notes how many dosage units were analyzed. The report must also specify the number analyzed and confirmed in the same manner as tablets and capsules.

If an analyst needs to combine samples from individual dosage units to obtain a conclusive identification, authorization must be obtained from the Technical Manager or designee, prior to the combination of samples. The authorization must be documented in the case file.

- F. When performing GC/MS analysis with a sodium hydroxide/hexanes extraction and there is not a controlled substance that can be identified, a second extraction has to be performed on the item. The analyst may perform the second extraction using methanol or sodium hydroxide/chloroform.
- G. If a liquid or residue is analyzed using sodium hydroxide/chloroform and codeine is identified, then a second extraction is required. Methanol or sodium bicarbonate/chloroform must be used. If the item is a tablet with a literary reference of codeine or a commercially labeled cough syrup with codeine, the second extraction is not required. Morphine is amphoteric and using sodium hydroxide/chloroform may cause the morphine to react with the NaOH and form codeine.
- H. If an item is analyzed using a methanol extraction and the chromatography is poor, i.e. peak fronting, no base-line separation, peak tailing, or no sharp peak formation, then a second extraction is required. The second extraction can be a methanol/sodium hydroxide/chloroform extraction of the original sample, or a second sample can be taken and extracted using sodium hydroxide/chloroform or sodium hydroxide/hexanes.

- I. If a case contains multiple items in one container, the item(s) tested must be given separate item numbers, different from the item(s) not tested. Internal packaging/containers do not have to have an item number or be on the items tab, but do need to be included on the report. Examples:
1. One plastic bag containing three spoons, and only one of the spoons is tested. The spoon tested would be labeled "Item 1A" and the remaining two spoons, not tested, would be labeled "Item 1B".
 2. One plastic bag containing multiple filter papers with residues, and only one of the filter papers is tested. Physically separate the filter paper(s) tested from filter papers not tested, by placing in separate appropriately labeled plastic bags.
"One plastic bag containing:
1A: One tightly wrapped bundle of filter paper containing white powder.
1B: Four tightly wrapped bundles of filter paper."
 3. One plastic bag containing a loose green leafy substance and three pipes. Physically separate the green leafy substance (to be tested) from other items tested or not tested, and label all appropriately.
"One plastic bag containing:
1A: A loose green leafy substance.
1B: One smoking device containing a residue.
1C: Two smoking devices containing residues.
Results:
1A: Cannabis....Schedule I.
1B: Tetrahydrocannabinols, Schedule III.
1C: No analysis."

If a case contains multiple items in one container and more than one item is going to be tested, the items may not be tested together or grouped into one extraction. Each item must be given separate item numbers. And they must be extracted and tested separately. It is not acceptable to combine two or more items into only one analysis. For example: One plastic tube containing 4 syringes. The syringes could be labeled as described with results as follows:

- "One plastic tube containing:
1A: One syringe with a residue.
1B: One syringe with a residue.
1C: Two syringes.
Results:
1A: Controlled substance1, Schedule XX.
1B: Controlled substance2, Schedule XX.
1C: No analysis."

This does not apply to tablets and capsules; refer to sections B, C and D regarding tablet labeling, testing and reporting.

- J. If a case contains items that are not analyzed and the items are considered “containers,” the contents need to be distinctly described. If the containers are empty, then describe them as such. Be clear in the description, so the reader understands what are contents and which are separate items.
1. Good Description example:
“One evidence envelope containing:
1A: One empty metal container on a ring and four plastic bags each containing residues
Results:
1A: No Analysis”
 2. Bad Description example:
“One evidence envelope containing:
1A: Four plastic bags each containing residues and one empty metal container on a ring
Results:
1A: No Analysis”
- K. Field Test Kits (FTK) submitted with evidence are NOT to be thrown away/disposed of. If received in the evidence container, the FTK needs to be placed in a heat-sealed bag to render it safe. The kit does not require an item number and should be included in the analysts notes and report or it can be combined with other non-analyzed items.. Examples:
- One evidence envelope containing Items 1A through 1B:
- 1A One plastic bag containing a green leafy substance...
- 1B One used marihuana field test kit and one package of papers
- or
- One evidence envelope containing one field test kit and:
- 1A One plastic bag containing a green leafy substance...
- 1B One plastic bag containing a crystal-like substance...
- L. On November 1st, 2014, Oklahoma legislature added trafficking levels to some prescription medications. The compounds and levels include:
- Morphine - 1,000 grams or more of a mixture
- Oxycodone - 400 grams or more of a mixture
- Hydrocodone - 3,750 grams or more of a mixture
- Benzodiazepine - 500 grams or more of a mixture

A definition or list of benzodiazepines is not included in the statute. The 2011 edition of DEA’s *Drugs of Abuse* online publication lists the following drugs as benzodiazepines: estazolam, flurazepam, temazepam, triazolam, midazolam, alprazolam, chlordiazepoxide, clorazepate, diazepam, halazepam, lorazepam, oxazepam, prazepam, quazepam, clonazepam, diazepam and clorazepate.

If 500 grams of any of these substances or any other benzodiazepines is encountered in casework, report the weight of the substance on the report and count or use the bench scale to get an approximate count to record in case notes.

7.4 Starting and Ending Date of Analysis:

The start date for analysis is the date the analyst assigns the case to themselves in the BEAST. The ending date of analysis is the date the analyst routes the report for technical review in the BEAST.

Unless otherwise indicated in the case notes, the date of sample weighing and sample extraction will be the date the analyst assigns the case to themselves in the BEAST. Also unless noted, the assigned analyst will be the analyst performing the weighing and extractions.

8. Abbreviation List

The following is a list of approved abbreviations for OSBI Controlled Substance Laboratory analysts' notes and within the chemistry matrix panel.

-	through	glr	green leafy residue
→	containing	glS	green leafy substance
~	approximately	gm	gram
apb	analyst provided bag	gpcr	glass pipe containing a residue
anw	approximate net weight	gsd	glass smoking device
app	approximate	gv	glass vial
blr	burnt leafy residue	gw	gross weight
bls	burnt leafy substance	h, hex	hexanes
br	burnt residue	hr	hand rolled
bps	brown paper sack	hrc	hand rolled cigarette
cap	capsule	hrCs	hand rolled cigarettes
cds	controlled dangerous substance	hspb	heat sealed plastic bag
cfp	coffee filter papers	l	labeled
cig	cigarette	lg	large
cigs	cigarettes	liq	liquid
clr	clear	lit ref	literary reference
cls	crystal-like substance	m	marked
con	containing	me	manila envelope
cont	containing	meoh	methanol
cpb	clear plastic bag	misc	miscible
cpzb	clear plastic ziploc bag	ml	milliliter
cs	crystalline substance	mls	milliliters
cw	cellophane wrapper	mnc, m-n-c, or m/n/c	methanol/sodium hydroxide/chloroform
czb	clear ziploc bag	msd	metal smoking device
d	diameter	na	no analysis
dl	modified duquenois levine	nc, n-c, or n/c	sodium hydroxide/chloroform
ea	each	net vol	net volume
ee	evidence envelope	nw	net weight
env	envelope	obn	Oklahoma Bureau of Narcotics

ext	extraction	ofw	off-white
fib mat	fibrous material	ow	off-white
g	gram	pb	plastic bag
gs	grams	pbc	plastic bag containing
pbhr	partially burnt, hand rolled	tab	tablet
peb	plastic "EVIDENCE" bag	tlc	thin-layer chromatography
pre	prescription bottle containing	tll	two-layer liquid
ps	powdery substance	tp	tan powder
psc	plastic sharps container	twb	tape wrapped bundle
psv	plastic sample vial	uom or UoM	uncertainty of measurement
pwb	plastic wrapped bundle	v, vol	volume
pzb	plastic zip lock bag	wbr	with a burned residue
qc	quality controls	we	white envelope
r	radius	wht	white
res	residue	wp	white powder
rls	rock-like substance	wps	white powdery substance
sm	small	w/	with
sj	sample jar	zpb	zip lock plastic bag
stu	safety tube		

9. Standard Forms for Documentation

The following is a list of forms that should be utilized for the maintenance documentation and recording Drug Reference Standards. DRQM forms are linked/embedded within this document; the other forms are linked/embedded within the Drug Protocols document.

Documentation Forms

<u>Maintenance Log Sheets for:</u>	<u>Other</u>
GC (DR30-1)	Drug Reference Standard Receipt and Usage Log (DRQM-4)
GC/MS (DR70-1)	
GC/MS/IRD (DR75-1)	
FTIR (DR60-1)	
Hydrogen Generator (DRQM-1)	
Air Compressors (DRQM-1)	

10. Documentation for the Training Manual

Included in the training manual are topics and checklists that should be used when training a new employee. Copies of the checklists that document an employee's training will be kept in the employee's training manual that is maintained by the unit's supervisor or the employee. Employees that have prior experience in drug chemistry may receive a modified training plan. The most recent training manual will be modified for such purposes; all modifications must be approved by the Forensic Chemistry Technical Manager. The most recent training manual is available in the Quality Management System (QMS).

11. Examination Documentation

- A. OSBI controlled substance analysts must store examination documentation within the image vault in the BEAST.
- B. Copies of the data/documentation may also be kept on the analyst's OSBI computer. The goal is to keep required documentation in the BEAST, without using excessive storage. If an analyst wants to keep multiple pages from a sample/shot, they may keep a copy on their OSBI computer and only upload the required documentation.
- C. Instrument documentation (GC, GC/MS, GC/IRD, or FTIR) that support the identification of substances must be kept in the image vault in the BEAST.
- D. *Rejected data:* Data that cannot be used to reach a conclusion because of quality issues or because it does not meet requirements of policy. Additional or alternative testing must be performed. Examples include but are not limited to: RI/RT not within the 2% window, standard information not on the scan, cocaine standard peak is not integrated, autotune was not run prior to casework, carryover in a sample or contamination of a blank.
- E. Rejected data must be entered on the rejected data panel in the BEAST matrix. The analyst must identify the data that is rejected (Instrument/Test and Validation Code if available), document the date of the rejection, and the reason for the rejection. Rejected data may be included in the final data uploaded to the image file, but each page must be marked "Rejected" and must still be listed on the rejected data panel.

For example:

1. If a GC is rejected due to QA/QC issues, the analyst would add the "Rejected Data Panel" to the matrix for the case and log the Instrument, item number, validation code, date, and reason.
2. If a GC/MS is rejected due to QA/QC issues, the analyst would add the "Rejected Data Panel" to the matrix for the case, log the Instrument, item numbers affected, validation code of the blank, date, and reason for rejection.

Example:

Case Methanol Blank has 58 in it and there were 3 tablets (items 1A, 1B and 1C) that followed. Analyst would enter Instrument, Items 1A-1C, the validation code of the blank which has the 58 ion present, date, and the reason on the "Rejected Data Panel."

3. If a GC/MS is rejected because it is outside the 2% retention index window, the analyst would add the "Rejected Data Panel" to the matrix for the case, log the instrument, item number, validation code for the scan and the blank (if the blank is not already included in the uploaded data file), date and reason for rejection.

4. If a TLC plate is run with three items from one case along with one item from a second case, and the entire plate changes color when the visualizing agent is applied. This is rejected data due to a quality issue (contamination of the solvent), so the analyst would add the "Rejected Data Panel" to the matrix for both cases and log the test, item numbers affected, date, and reason.
- F. Samples/shots that are not used in the identification, but not rejected, will be uploaded into the image vault of the BEAST. The analyst must keep at least one page from each sample/shot.

Examples include but are not limited to:

1. A sample must be re-run to increase/decrease concentration on GC/MS because there are not enough ions, too many ions, or the ratio of the ions are off (but the retention index is within 2%). For example: Drug100 is too weak, reshoot using Drug50. Analyst would keep both the Drug100 and the Drug50 scans in the uploaded data file.
 2. An extraction needs to be performed because chromatography is poor. For example: Extract initially into methanol, but chromatography is poor, so extract into MeOH/NaOH/CHCl₃. Analyst would keep both scans in the uploaded data file
 3. FTIR scan indicates a mixture
- G. Autotunes and hydrocarbon ladder data from the GC/MS do not have to be uploaded into the case file. Copies of instrumental data, QA/QC (autotunes, ladders, blanks & cocaine) and PDFs are kept on each instrument computer and are backed up routinely.
- H. For cases in which weights are included in the results section of the examination report, a PDF document or Excel spreadsheet with calculations for the total weight and uncertainty is to be included in the case file. This is only required for cases in which two or more items are involved in the calculations. The items, weights, units, and uncertainty must be labeled and the totals identifiable. The OSBI Lab number, analyst name and date (of calculation) must also be documented on the document.
- I. A complete step by step detailed description of the PDF procedure is available in DR-50. This procedure should be followed by all controlled substance analysts when using PDF exam documentation.
- J. Examination documentation maintained in the BEAST image vault is not required to have page numbers. Any examination documentation printed and maintained as a permanent record in the case file are required to follow all requirements listed in the Criminalistics Services Division Laboratory Quality Manual.

12. Case Reviews

QP-31 requires that all OSBI case files be reviewed by an individual other than the reporting analyst prior to being released to an outside agency. Administrative and Technical Reviews will meet all requirements set forth by QP-31 and will include a thorough review of the entire case file and Criminalistics Examination Report by a qualified reviewer. The specific procedure followed by the controlled substances laboratory will be as follows:

- A. Once the analyst has signed the case in the BEAST, the case will be distributed for review utilizing the routing function in the BEAST. Analysts will rotate reviews through all approved drug analysts before sending another review to the same analyst. Exceptions to this can be made for rush cases or if the next reviewing analyst has notified drug analysts of vacation or extended leave
- B. If any changes or corrections are required, QP-31 requires the case to be routed for correction. The reviewer must include an explanation in the comments field of the routing box describing what must be corrected. Once corrections have been made, the case will be routed back, with corrections/changes documented in the comments field, to the original reviewing analyst for additional review.
1. If the correction involves multiple cases, the analyst must notify the Technical Manager, or designee, of all cases that were reviewed for necessary correction and cases in which corrections were required and made.
 2. If the two analysts are unable to come to an agreement, the Forensic Chemistry Technical Manager should be consulted and the case routed to the Technical Manager, or designee, using the "Route to Technical Manager" code. The Technical Manager, or designee, will then route the case back with their decision; the Technical Review will be completed by the original reviewer.
 3. The reviewing analyst is considered a "designee" approved by the Technical Manager for approving Simple Corrections. At any time during the review process, if the reviewing analyst has a question as regarding the Correction they need to contact the Technical Manager for clarification. The Technical Manager receives a daily report of all routings and will document any relevant corrections on the Simple Correction Log. The Technical Manager maintains authority to override decisions of the reviewing analyst.
- C. By answering "yes" to the statement on the Technical Review form the reviewer agrees to the following statements.

For all cases:

1. Tests conform with the appropriate Controlled Substances Unit protocols and methods, the Controlled Substances Quality Assurance Manual and CSD policies and procedures.
2. Results, opinions and interpretations are accurate, properly qualified and supported by the technical records.
3. Spectra, data, images, and observations, etc. that support findings, are included in the file.
4. Conclusions reported are in accordance with conclusions listed in the analytical notes.
5. Result reporting is consistent with protocol recommended reporting.
6. QA/QC documentation is included (blanks, standards, etc.).
7. Handwritten notes, calculations, analyst notations in PDFs and the naming of PDFs have been reviewed and compared to documentation in the BEAST Matrix.

For cases in which instrumentation was used:

8. Case number and analyst initials/name are on each page of their notes.

For cases in which GC and GC/MS were used:

9. GC retention time differences are 2% or less from the standard.
10. GC/M retention index differences are 2% or less from the standard.

For cases in which FTIR was used:

11. FTIR ValPRO System Qualification of the accessory is performed, or the spectra of a polystyrene standard obtained, on the day of casework.
12. FTIR background is performed before each sample.

For cases in which weights are included in the results:

13. PDF document/page or Excel spreadsheet, with calculations of weight and uncertainty, is included in the case file, for cases in which the total weight consists of two or more items.
14. Totals/calculations are correct.
15. Uncertainty used for calculations is the current/correct uncertainty for the reporting laboratory.
16. The coverage probability statement is on the report.

- D. Approving the report will be documentation that an administrative review was completed by the approving analyst. Once a technical/administrative review is complete, the reviewing analyst will notify the requesting analyst that the review has been completed. If reports must be mailed, they should be sent from the originating laboratory.

Approval

Technical Manager Heather Schafstall
Heather Schafstall

Date: 07-19-18

Laboratory Director Andrea Fielding
Andrea Fielding

Date: _____

History

Rev. #	Issue Date	History
0	08-28-00	Original Issue
0	06-20-01	Review with no changes.
1	09-18-02	Change in format and wording
2	11-01-03	Annual Review (07-23-03). Clarification of acceptable extraction blanks in section 3.4.7.
3	12-01-04	Annual Review (8-03-04). Addition of annual requirement to change split vent line and split vent trap on instruments. Reformatted quality manual.
4	07-01-05	Annual Review (05-26-05). Addition of abbreviation list.
5	12-15-06	Annual Review (09-27-06). Correction's to abbreviation list.
6	12-31-07	Annual Review (9-21-07). Addition of Raman IR. Typographical Corrections.
7	12-31-08	Annual Review (10-08-08). Revision of accepted abbreviations for use in BEAST. Addition of ability to use BEAST chemical inventory program. Addition of blank methods to run at consistent split ratios and runtimes as drug analysis methods.
8	05-15-09	Updated abbreviation list.
9	07-31-09	Addition of PDF option for drug analysts. Addition of approval line for Division Director. Addition of distribution footer.
10	12-31-09	Annual Review (09-11-09). Change laboratory director. Addition of split vent line and filter counts. Addition of tablet reporting and labeling requirements. Updated PDF document. Added distribution to header information. Removed analysis worksheets. Addition of technical review requirements.
11	02-15-10	Change of document location for PDF document and training manual to comply with McAlester ISO audit findings.
12	12-31-10	Annual Review (9-29-10). Added function testing requirement for replacement or repaired equipment (1.4-A). Deviation procedure changed to follow Laboratory Quality Manual QP3. Section 4 changed to incorporate drug reference standard procedure formerly in QP 26.
13	01-12-11	Added administrative review procedure to meet requirements of QP-31. Updated attached forms to have designator "AΩ" to indicate a one-page form.
14	12-20-11	Annual Review (09-15-11). Addition of background subtractions to cases requiring technical review. Sampling of tablet(s) and capsule(s) limited to ½ of each tablet(s) or capsule(s). GCMS solvent blank requirements moved to DR-70.
15	12-20-12	Annual Review (04-03-12). Incorporated deviation issued 04-09-12 requiring 100% technical review. Incorporated deviation issued 04-09-12 designating the start and ending dates of analysis. Incorporated deviation issued 04-23-12 adding OBN (Oklahoma Bureau of Narcotics) to the approved abbreviation list

16	12-20-13	<p>Annual Review (11-11-13). Changed references to “drug identification” to “controlled substance” (where appropriate) and references to “Drug Technical Manager” or “Technical Manager” to “Forensic Chemistry Technical Manager.” Renumbering of sections 5-12. Change of Technical Manager.</p> <p>Section 1: Removed references to HPLC, Gas Monitors and XRFS. Added changing split vent filter for Agilent GC 7890. Replaced “Hewlett Packard” with “ASAP Analytical” in Section 1.4.A. Added split vent filter to Agilent 7890 GC to Section 1.4.B & G. 1.4.G – modified chart, removed IRD mirror alignment, added Protocol Reference, 1.4.H-Changed Refrigeration Monitoring forms, from DRQM-2 to OSBI CSD QPA 6.4.1, and referred to QP 6.3 for monitoring refrigerators containing evidence. Updated all the Standardized DRQM-5 forms to reflect changes (Worksheet name, addition of tables, revision date). DRQM-2 has been rescinded.</p> <p>Section 2: 2.4.A-Changed references to Chemical Formulation Worksheet (QP8.1) to Reagent Formulation Worksheet (DRQM-5), and removed location of form QP8.1. Added initials of receiving analyst to be marked on container to 2.4-D.</p> <p>Section 4: Changed 4.3 from designated Criminalist Administrator to Criminalistics Division Director. 4.6- Removed notifying Administrator upon receipt of Schedule I or II. Changed Section 4.15 to reference #5 instead of #4. Added Initials and date to 4.7 and 4.9, Standards.</p> <p>Section 5: Moved section 7 to 5. Clarified Standards and Reagent sections. Removed instrument maintenance chart (duplicate of section 1.4.G).</p> <p>Section 6: Added Weight Reference Standards as a new section.</p> <p>Section 7: 7.3.A-Removed HPLC from Non-Confirmatory Tests. Added requirement for GC prior to analysis by GC/MS or GC/IRD. 7.3.C - Added “and confirmed” to the number of tablets or capsules analyzed. Labeled second paragraph of Section F as Section G. 7.3.G – Added sodium/hexanes as second extraction. Added Section H, regarding multiple items in one container.</p> <p>Section 8: Removed reference to CSD QM for QP8.1. Clarified location of forms on Intranet and OSBI network. Added reference to Reagent Formulation Worksheet (DRQM-5). Added E/P Standard Ladder and 10% Acetic Acid to Reagent Formulation Worksheets.</p> <p>Section 9: Incorporated modified training for new employees with prior experience.</p>
17	12-20-14	<p>Section 1.4.G Added GC/IRD to Gold Seals, added “whichever comes first” to Deionizer bag replacement.</p> <p>Section 1.4.H Changed reference QP 6.3 to PQ 6.4</p> <p>Section 4. Removed “log book” and added option for scanning and storing on network for destroyed or expended drug standards. Added storage of initial log sheet and verification data in log book or electronically on network. Added testing requirements for creating secondary standards.</p>

		<p>Added 4.17 regarding inventory and analyzing and documenting secondary standards</p> <p>Section 6.2 Added statement regarding traceability and uncertainty for weights and approximate volumes.</p> <p>Section 7.3 Removed Liebermann’s and Sodium Nitroprusside from Presumptive Tests.</p> <p>Section 8 Added sodium hydroxide/chloroform and hexanes</p> <p>Section 9 Removed the server name for the Contsub\Forms folder. Added Acetonitrile and removed Liebermann’s Reagent, Sodium Nitroprusside Reagent 1 and 2.</p> <p>Section 11 Added paragraph regarding instrument data, storage of instrument and raw data.</p>
18	12-31-15	<p>Annual Review (11-30-15) Section 4.8 Added network folder for verification documentation.</p> <p>Section 4.10 Changed “section” to network folder for expended documentation when transferring standard.</p> <p>Section 7.3 Added wording from deviation as Section G and relabeled remaining sections of 7.3.</p> <p>Section 8 Added mnc-methanol/sodium hydroxide/chloroform and uom-uncertainty of measurement, d-diameter, r-radius, v and vol-volume, meoh-methanol</p> <p>Section 9 Changed Refrigerator Temp Log to Temperature Monitoring Form, QPA 6.4.1.</p> <p>Sections 12 & 13 Removed requirements for routing “Review Finalized and Approved” and other forms of notification after review is completed.</p> <p>Section 13 Added instructions on routing to TM in case of disagreement.</p>
19	12-31-16	<p>Annual Review (11-01-16) Added analyst responsibility for clarification statement directly before the Objective</p> <p>Section 1.4.B Added sentence regarding not changing the glass wool, without changing the liner/septa.</p> <p>Section 1.4.F. Added scale calibration statement.</p> <p>Section 2.4.E Added the lot number will be the date the reagent is made.</p> <p>Section 7.3 Added new language to section E for labeling and sampling of evidence in the form of dosage units and relabeled section E through I to F through J.</p> <p>7.3.E (now section F) Removed adding analyst initials requirement on TIC for manual scan. Added language clarifying the requirements for documentation when a substance cannot be identified.</p> <p>Section 11 Added requirement for PDF or Excel spreadsheet for totals for weights and uncertainties in where the weights are reported in the results.</p> <p>Section 12 Removed automatic notification that Admin Review was completed via “Approve” button.</p> <p>Section 13 Incorporated the wording for the Technical Review form. Added wording for trafficking cases and uncertainty. Added requirement for PDF</p>

		or Excel spreadsheet for trafficking cases in which 2 or more items are involved. Removed automatic notification that review was completed via “Approve” button.
20	12-31-17	<p>Annual Review. (10-31-17) Reformatting: added Table of Contents, moved Objective, forms were removed from “end of the manual”, combined History into document</p> <p><u>Section 1.4.A</u> Instrument logs are incorporated into the Protocol Manual</p> <p><u>Section 1.4.C</u> Added storage of maintenance log in electronic format.</p> <p><u>Section 1.4.G</u> Added DR-75 and GC/IRD to the maintenance schedule.</p> <p><u>Section 2.2.B</u> Added archiving of historical reagent information and reagent information will be kept in ChemInventory.</p> <p><u>Section 2.4.A</u> Added reagent information to be documented in ChemInventory. Uncontrolled DRQM-5.</p> <p><u>Section 2.4.B</u> Removed requirement for reagent logbook and added labeling requirement for prepared reagents.</p> <p><u>Section 2.4.C</u> Removed requirement for formulation sheets and distribution if used. Added “if used” requirement for formulation sheets will be uploaded.</p> <p><u>Section 2.4.E</u> Original E paragraph renumbered to F. Added function verification requirements to new E.</p> <p><u>Section 3.2.B</u> Added archiving of historical solvent information and solvent information will be kept in ChemInventory.</p> <p><u>Section 3.4.C</u> Reformatted paragraph</p> <p><u>Section 3.4.D</u> Original Section D was renumbered to E. Added section on ChemInventory documentation requirements.</p> <p><u>Sections 3.4.E, F and G</u> Removed requirement for storing solvent logs in notebooks. Added (to new Section E) that GCMS scans will be stored in ChemInventory. Deleted old sections E and F and uncontrolled DRQM-3 and added “if used” requirement for uploading the form. Renumbered G to F.</p> <p><u>Section 4</u> Changed numbering to letters</p> <p><u>Section 5</u> Added “storage requirements (as applicable)” to Reagents</p> <p><u>Section 6.2</u> Changed numbering to letters</p> <p><u>Section 6.2.A</u> Added statement regarding weights as being critical and vendors will be evaluated.</p> <p><u>Section 7.3.A</u> Added “(LSD)” to Ehrlich’s. Added GC/IRD for identification of THC.</p> <p><u>Section 7.3.B</u> Moved statement regarding analysis of pharmaceuticals with no label to DR-5</p> <p><u>Section 7.3.C</u> Reformatted paragraph. Added tablet fragments. Removed statement regarding one tablet being analyzed in most cases. Added requirement that multiple samples of tablet/capsule/etc. must be analyzed individually.</p> <p><u>Section 7.3.C.3</u> Changed from single tablet cases to cases containing single tablet(s).</p>

		<p><u>Section 7.3.D</u> Added tablet fragments</p> <p><u>Section 7.3.E</u> Reworded report requirement sentence for dosage units.</p> <p><u>Section 7.3.F</u> Moved Paragraph F to DR-70. Renumbered remaining paragraphs</p> <p><u>Section 7.3.I</u> Added internal packaging do not have to be numbered but do need to be on the report.</p> <p><u>Sections 7.3.J, K and L</u> New paragraphs</p> <p><u>Section 8</u> Added abbreviations: dl, qc, tlc</p> <p><u>Section 9</u> Uncontrolled DRQM-3 and DRQM-5, removed section of paragraph regarding those forms. Added where controlled forms can be found. Removed list of Solvent Log Sheets and Reagent Formulation Worksheets.</p> <p><u>Section 10</u> Changed location of training manual from Intranet to QMS</p> <p><u>Section 11</u> Added numbering/lettering. Removed references to hard copy files. Changed and added documentation requirements. Added rejection of data definition and requirements/examples. Added specific requirements on UoM casefile documentation. Removed PDF attachment requirements that are addressed in CSD QM.</p> <p><u>Section 12</u> Changed numbering to lettering. Changed name of section. Reworded the first paragraph to merge sections 12 & 13. Added requirement to notify TM of multiple cases involved in a correction.</p> <p><u>Section 12.C.1</u> Added CSD policies and procedures.</p> <p><u>Section 13</u> Added numbering/lettering. Changed wording to TR statements 1 & 2 and added 7, to meet ANAB requirements. Added corrections/changes documented in comments field in paragraph D. Merged with Section 12 and removed redundant wording. TR requirement 1. added “and CSD policies and procedures.”</p>
21	07-20-18	<p><u>Section 4.H</u> Reformatted paragraph into subsections. Added requirements in subsection 1 regarding unverified standards and the vendor.</p> <p><u>Section 6.2.A</u> Changed wording for outside vendor’s accreditation requirements.</p> <p><u>Section 6.2.E</u> Changed wording for outside vendor’s accreditation requirements.</p> <p><u>Section 7.3.C</u> Moved subsection 2 & 3 to 3 & 4 and added a new subsection 2 regarding multiple tablets of one tablet type and the tablet statement requirement at the bottom of the report.</p> <p><u>Section 7.3.K</u> Incorporated deviation regarding FTK</p> <p><u>Section 7.4</u> Added second paragraph regarding date of sampling and extractions.</p> <p><u>Section 12.B</u> Added subsection 3 regarding analysts as designees of Technical Manager for Simple Corrections.</p>